

## WATER OR AIR FLOW

### Air Flow

For flow standards or other critical uses, O'Keefe Controls Co. can provide measured air flow data for any of the standard precision orifice products. This optional service is available at a nominal charge and can be provided for pressure or vacuum flow conditions in the standard ranges listed below. A calibration sheet with measured data is provided for each orifice (see sample below).

Data is obtained using a digital mass flowmeter traceable to the National Institute of Standards Technology (NIST). Standard conditions are 70°F and 14.7 psia (See photo page 45).

Air flow can be measured for any of the following conditions:

- Pressure 0-5, 0-10, 0-20, 0-50, 0-100 psig
- Vacuum 0-30" Hg
- Flow 0-.002, 0-.01, 0-.02, 0-.05, 0-.1, 0-.2, 0-.5, 0-1, 0-2, 0-5, 0-10, 0-20, 0-50, 0-200 SLPM
- Flow 0-.004, 0-.02, 0-.04, 0-.10, 0-.2, 0-.4, 0-1, 0-2, 0-4, 0-10, 0-20, 0-40, 0-100, 0-400 SCFH



#### *C<sub>v</sub> Measurement*

*Water at a constant height equivalent to 1 psig is used to determine the amount of water that passes through a test orifice in a measured time period. The C<sub>v</sub> is calculated from the measured data.*

### Water Flow


Calibration of standard orifices using water flow is accomplished by measuring the quantity of water flowing through an orifice in a given time period at a specific inlet pressure.

The calibration procedure determines the water flow rate through the orifice and the corresponding C<sub>v</sub> of the orifice is calculated. The C<sub>v</sub> can then be used to predict the water flow rate at varying differential pressure conditions using the equation:

$$Q = C_v \sqrt{\Delta P}$$

For liquids other than water flow rate can also be predicted using the C<sub>v</sub> method. See page 47 for details on liquid flow - C<sub>v</sub> method.


A calibration sheet with measured data is provided for each orifice. The calculated C<sub>v</sub> is the average value computed from five separate tests (see sample below).



### MEASURED DATA POINTS

#### Air Flow

Pressure Units -	Air Flow Units -	Barometric Pressure _____
_____	_____	Ambient Temperature _____
_____	_____	Flowmeter _____
_____	_____	OKC Part No. _____
_____	_____	I/D No. _____
_____	_____	Purchase Order _____
_____	_____	Customer _____
_____	_____	Customer P/N _____
_____	_____	Invoice _____
_____	_____	Data By _____
_____	_____	Date _____



### C<sub>v</sub> TEST CALIBRATION

#### Water Flow

Test	Weight of Collected Water	Fill Time (min)	Calculated C <sub>v</sub>
1			
2			
3			
4			
5			
Average C <sub>v</sub>			

O'Keefe Part Number \_\_\_\_\_ Customer Name \_\_\_\_\_

I/D Number \_\_\_\_\_ Customer Purchase Order \_\_\_\_\_

O'Keefe Invoice No. \_\_\_\_\_ Customer Part Number \_\_\_\_\_

Barometric Pressure \_\_\_\_\_ Temperature \_\_\_\_\_

Date \_\_\_\_\_ Data By \_\_\_\_\_